



Part No.: H-020-000798-00

MGIEasy

Magnetic Beads Genomic DNA Extraction Kit

Instructions for Use

Version: 1.0

Leading Life Science Innovation

Address 1: Building 24, Stage 3.1, BioLake Accelerator, No.388,
2nd Gaoxin Road, East Lake High-Tech Development
Zone, 430075, Wuhan, P.R.China

Address 2: Building B13, No.818, Gaoxin Avenue, East Lake
High-Tech Development Zone, 430075, Wuhan,
P.R.China

E-mail: MGI-service@mgi-tech.com

Website: en.mgi-tech.com

Research Use
Only

Wuhan MGI Tech Co., Ltd.

About the instructions for use

This instructions for use is applicable to MGIEasy Magnetic Beads Genomic DNA Extraction Kit. The version of the instructions for use is 1.0 and the kit version is 1.0.

This instructions for use and the information contained herein are proprietary to Wuhan MGI Tech Co., Ltd. (hereinafter referred to as MGI), and are intended solely for the contractual use of its customers in the use of the products described herein and for no other purpose. This instructions for use and its contents shall not be reprinted, reproduced, modified, distributed, or disclosed to others, in whole or in part, without prior written consent from MGI.

MGI makes no commitment to this instructions for use, including (but not limited to) any special commercial purpose and any reasonable implied warranties. MGI has taken measures to ensure the correctness of this instructions for use. However, MGI is not responsible for any missing parts of the instructions for use and reserves the right to revise the instructions for use and modify the kit, so as to improve the reliability, performance or design.

Figures in this instructions for use are for illustrative purpose only. The content may be slightly different from the set. For the most up-to-date details, refer to the kit purchased.

MGIEasy™ is the trademark of MGI, or its subsidiaries in China and/or other countries. Other names and trademarks mentioned in this instructions for use may be the property of their respective companies and subsidiaries.

©2023 Wuhan MGI Tech Co., Ltd. All rights reserved.

Revision history

Version	Date	Description
1.0	August 15, 2023	Initial release

Contents

Chapter 1 Introduction	1
1.1 Product name	1
1.2 Specifications	1
1.3 Intended use	1
1.4 Working principle	1
1.5 Main components	2
Chapter 2 Applicable device	3
Chapter 3 Sample requirements	3
3.1 Applicable sample	3
3.2 Sample amount requirements	4
3.3 Sample storage	4
3.4 Sample transportation	4
3.5 Sample safety	5
Chapter 4 Operation	5
4.1 Preparing materials	5
4.2 Pretreating samples	6
4.3 Extracting the nucleic acids	8
Chapter 5 Warnings and precautions	13
Appendix 1 Manufacturer information	14

Chapter 1 Introduction

1.1 Product name

MGIEasy Magnetic Beads Genomic DNA Extraction Kit

1.2 Specifications

Kit name	Model	Cat. No.	Specification
MGIEasy Magnetic Beads Genomic DNA Extraction Kit	WDT-96	940-000972-00	96 Preps
MGIEasy Magnetic Beads Genomic DNA Extraction Kit	WDT-864	940-000973-00	864 Preps

1.3 Intended use

This set is used to extract, enrich and purify nucleic acids.

1.4 Working principle

By using the unique, high-binding, super-paramagnetic beads, this kit is used to extract high-quality genomic DNA quickly and easily from blood, saliva stored by MGI saliva DNA collection kit, fresh saliva, buccal swabs, animal tissues, cells and other samples. The extracted genomic DNA can be used for various routine applications, including enzyme digestion, PCR, real-time PCR, library preparation, chip hybridization and high-throughput sequencing.

1.5 Main components



- Tips**
- Do not mixedly use reagents from different batches of kits.
 - Store the kit in a dry environment. To store Proteinase K and Magnetic Beads H for a longer time, store these two reagents in a refrigerator at 2 °C to 8 °C .
 - That precipitation forms in Buffer LB and Buffer W1 is normal and does not affect the reagent performance. Before use, preheat the reagents for 10 minutes in a water bath at 37 °C to dissolve the precipitation and mix the reagents thoroughly.
 - Before use, take out all components in the reagent kit, equilibrate to room temperature (10 °C to 30 °C) and mix them thoroughly before adding to wells.
 - Buffer EB consists of 10 mM Tris-HCl (pH 8.0) and 0.5 mM EDTA (pH 8.0). Please prepare the elution buffer according to your specific needs.

Table 1 MGIEasy Magnetic Beads Genomic DNA Extraction Kit (WDT-96)
Cat. No.: 940-000972-00

Name	Component	Specification	Storage condition	Validity period	Transportation condition
MGIEasy Magnetic Beads Genomic DNA Extraction Kit Cat. No.: 940-000972-00	Buffer LS	22 mL/bottle×1	2 °C to 30 °C	18 months	2 °C to 30 °C
	Buffer LB	30 mL/bottle×1			
	Buffer W1	28 mL/bottle×1			
	Buffer W2	30 mL/bottle×1			
	Buffer EB	20 mL/bottle×1			
	Proteinase K	2 mL/tube×1			
	Magnetic Beads H	2 mL/tube×1			

Table 2 MGIEasy Magnetic Beads Genomic DNA Extraction Kit (WDT-864)
Cat. No.: 940-000973-00

Name	Component	Specification	Storage condition	Validity period	Transportation condition
MGIEasy Magnetic Beads Genomic DNA Extraction Kit Cat. No.: 940-000973-00	Buffer LS	200 mL/bottle×1	2 °C to 30 °C	18 months	2 °C to 30 °C
	Buffer LB	260 mL/bottle×1			
	Buffer W1	240 mL/bottle×1			
	Buffer W2	140 mL/bottle×2			
	Buffer EB	180 mL/bottle×1			
	Proteinase K	18 mL/tube×1			
	Magnetic Beads H	18 mL/tube×1			

Chapter 2 Applicable device

MGISP-NE384RS Automated Nucleic Acid Extractor

Chapter 3 Sample requirements

3.1 Applicable sample

This product is applicable to blood, saliva stored by MGI saliva sample collection kit, fresh saliva, buccal swabs, amniotic fluid, cells and animal tissues.

3.2 Sample amount requirements

Sample type		Sample amount	
		Manual extraction	Extraction on MGISP-NE384RS
Blood	Fresh/frozen blood	200 µL	200 µL
	Anticoagulant blood of poultry, birds, amphibians, or lower organisms	5 µL to 10 µL	5 µL to 10 µL
Saliva	Saliva/buccal swab stored by MGI saliva sample collection kit	500 µL	500 µL
	Fresh saliva	200 µL	200 µL
Cell		$\leq 5 \times 10^6$	$\leq 5 \times 10^6$
Amniotic fluid		3 mL to 5 mL	3 mL to 5 mL
Animal tissue		2 mg to 50 mg	5 mg to 15 mg

3.3 Sample storage

- For samples of blood, amniotic fluid, cell and animal tissue that could be tested within 24 hours, store them at 2 °C to 8 °C . For those that could not be tested within 24 hours, store them at -70 °C or below, or in a freezer at -25 °C to -15 °C . During storage, do not freeze and thaw samples frequently.
- For the fresh saliva sample, use it immediately after collection. It is recommended to use the MGI saliva sample collection kit (MGI, Cat. No.: 940-001262-00/1000025954) to collect saliva samples which then could be stored at room temperature.
- Do not freeze and thaw frozen samples frequently. Otherwise, the DNA quality may decrease.

3.4 Sample transportation

- For samples of blood, amniotic fluid, cell and animal tissue, use the dry ice for transportation for up to 7 days. During transportation, avoid frequent freeze-thaw cycles.
- For samples stored by MGI saliva sample collection kit, transport them at room temperature.

3.5 Sample safety

All samples are regarded potentially infectious. All samples should be extracted after being inactivated according to relevant national regulations.

Chapter 4 Operation

4.1 Preparing materials

Prepare the following materials:

Table 3 User-supplied materials

Type	Item	Description
Equipment	MGISP-NE384RS automated nucleic acid extractor	<ul style="list-style-type: none"> • MGI, Cat. No.: 900-000357-00 • For use in automated extraction
	Mini centrifuge	With a speed no less than 12000 rpm
	Vortex mixer	None
	Thermomixer compact	It can be replaced by a water bath
	1.5 mL magnetic rack	None
	Pipette	1 mL/200 µL/20 µL/10 µL
Reagent	Absolute ethanol	Analytically pure
	Isopropanol	Analytically pure
	RNase A	<ul style="list-style-type: none"> • 20 mg/mL • DNase-free
Consumables	Saliva sample collection kit	MGI, Cat. No.: 940-001262-00 MGI, Cat. No.: 1000025954
	Tips	1 mL/200 µL/20 µL/10 µL
	Centrifuge tube	<ul style="list-style-type: none"> • 5 mL/1.5 mL • DNase-free and RNase-free

4.2 Pretreating samples

Pretreat samples according to different types of samples. In automated extraction, there is no need to pretreat blood and saliva samples.

 **Tips** Please thaw and mix the frozen samples thoroughly before use.

4.2.1 Blood sample


Perform the following steps:


1. Perform different steps according to the blood type.
 - For fresh or frozen blood, add 200 μL of sample into a new 1.5 mL centrifuge tube.
 - For anticoagulant blood of poultry, birds, amphibians, or lower organisms, add 5 μL to 10 μL of sample into a new 1.5 mL centrifuge tube.
2. Add Buffer LS into the tube to bring the final volume to 200 μL if the total volume is less than 200 μL .
3. Add 20 μL of Proteinase K into the tube, and mix the tube thoroughly by vortexing.
4. Add 300 μL of Buffer LB into the tube, and mix the tube thoroughly by vortexing.

4.2.2 Saliva sample

Perform the following steps:

1. Perform different steps according to the saliva type.
 - For saliva or buccal swab stored by MGI saliva sample collection kit, add 500 μL of sample into a new 1.5 mL centrifuge tube.

 **Tips** Mix the sample thoroughly before sampling to obtain sufficient DNA sample.
 - For fresh saliva sample, add 200 μL of sample into a new 1.5 mL centrifuge tube.

 **Tips** If the sample is not extracted immediately, add 300 μL of Buffer LB into the tube, mix thoroughly and store it at room temperature. Perform the next step within 24 hours.
2. Add 20 μL of Proteinase K into the tube, and mix the tube thoroughly by vortexing.
3. Perform different steps according to the saliva type.

- For saliva or buccal swab stored by MGI saliva sample collection kit, or fresh saliva stored by Buffer LB, go to the next step directly.
- For fresh saliva sample, add 300 μ L of Buffer LB into the centrifuge tube and mix the tube thoroughly by vortexing.

4.2.3 Cell sample

Perform the following steps:

1. Add cell suspension sample whose extraction volume does not exceed 5×10^6 into a new 1.5 mL centrifuge tube.
 - For cell suspension sample with high concentration, add Buffer LS to dilute the sample to that of less than 5×10^6 cells/mL.
 - For adherent cells, perform the following steps:
 - a. Prepare cell suspension from sample. Add 1 mL of sample into a new 1.5 mL centrifuge tube.
 - b. Centrifuge the tube in a centrifuge at 10000 rpm for 1 minute.
 - c. Remove the supernatant, add 200 μ L of Buffer LS into the tube, and vortex it to suspend it completely.
2. Add 20 μ L of Proteinase K into the tube, and mix the tube thoroughly by vortexing.
3. Place the tube into a thermomixer compact to incubate it for 30 to 60 minutes at 65 °C with a speed of 1000 rpm to 1200 rpm. Briefly centrifuge the tube when the cell is dissolved.
4. Add 300 μ L of Buffer LB into the centrifuge tube and mix the tube thoroughly by vortexing.

4.2.4 Amniotic fluid sample

Perform the following steps:

1. Add 3 mL to 5 mL of amniotic fluid sample into a new 5 mL centrifuge tube.
2. Centrifuge the tube in a centrifuge at 6000 rpm for 2 minutes.
3. Remove the supernatant without aspirating the pellet.
4. Add Buffer LS into the tube to bring the final volume to 200 μ L. Transfer the reagent in the tube to a new 1.5 mL centrifuge tube.
5. Add 20 μ L of Proteinase K into the tube, and mix the tube thoroughly by vortexing.

6. Place the tube into a thermomixer compact to incubate it for 30 to 60 minutes at 65 °C with a speed of 1000 rpm to 1200 rpm. Briefly centrifuge the tube when the amniotic fluid sample is dissolved.
7. Add 300 µL of Buffer LB into the centrifuge tube and mix the tube thoroughly by vortexing.

4.2.5 Animal tissue sample

Perform the following steps:

1. Prepare 2 mg to 50 mg of fresh or frozen tissue sample, use a surgical knife or a pair of surgical scissors to cut the sample as big as a sesame seed and then add them into a new 1.5 mL centrifuge tube.
2. Add 200 µL of Buffer LS into the tube, and vortex it to suspend it completely.
3. Add 20 µL of Proteinase K into the tube, and mix the tube thoroughly by vortexing.
4. Place the tube into a thermomixer compact to incubate it for 30 to 60 minutes at 65 °C with a speed of 1000 rpm to 1200 rpm. Briefly centrifuge the tube when the tissue is dissolved.
5. Add 300 µL of Buffer LB into the centrifuge tube and mix the tube thoroughly by vortexing.

4.3 Extracting the nucleic acids



Tips You can extract the nucleic acids manually or on automation devices. For automated nucleic acid extraction, ensure that you have prepared applicable consumables.

4.3.1 Extracting the nucleic acids manually

Before extraction, add absolute ethanol into Buffer W1 and Buffer W2 according to the label.

Perform the following steps:

1. Place the tube into a thermomixer compact to incubate it for 15 minutes at 65 °C with a speed of 1000 rpm to 2000 rpm.
2. Add 350 µL of isopropanol into the tube and mix the tube thoroughly by vortexing. It is normal that the flocculent precipitation forms in the tube.
3. Add 20 µL of Magnetic Beads H into the tube and mix the tube thoroughly by vortexing. Place the tube at room temperature for 5 minutes during which vortex it every 2 minutes.



Tips Mix Magnetic Beads H thoroughly by vortexing before use to ensure that beads resuspend completely.

4. Centrifuge the tube briefly and place it on the magnetic rack for 2 minutes. When Magnetic Beads H is adsorbed completely, use a pipette to slowly remove the supernatant.
5. Remove the tube from the rack. Add 500 μ L of Buffer W1 and mix the tube thoroughly by vortexing for 1 to 2 minutes.



Tips Please do mix the tube thoroughly. Otherwise, the DNA purity may be affected.

6. Place the tube on the magnetic rack for 1 minute. When Magnetic Beads H is adsorbed completely, use a pipette to slowly remove the supernatant.
7. Remove the tube from the rack. Add 600 μ L of Buffer W2 and mix the tube thoroughly by vortexing for 1 to 2 minutes.
8. Place the tube on the magnetic rack for 1 minute. When Magnetic Beads H is adsorbed completely, use a pipette to slowly remove the supernatant.
9. Repeat step 7 to 8 once. During operation, remove the remaining liquid in the tube as much as possible.
10. Place the tube on the magnetic rack, and decap and dry the tube at room temperature for 5 to 10 minutes to ensure that the ethanol evaporates completely.
11. Remove the tube from the rack. Add 50 μ L to 100 μ L of Buffer EB into the tube and mix the tube thoroughly by vortexing. Place the tube into a thermomixer compact to incubate it for 5 minutes at 56 °C with a speed of 1000 rpm.



Tips To remove RNA, add Buffer EB containing RNase A into the tube at this step. It is recommended to add 0.5 μ L of RNase A (20 mg/mL) for every 100 μ L of Buffer EB.

12. Place the tube on the magnetic rack. When Magnetic Beads H is adsorbed completely, transfer the DNA solution into a new 1.5 mL centrifuge tube. Mark the tube and store it at -20 °C or below.



Tips

- The recommended volume of Buffer EB is no less than 50 μ L.
- When DNA yield is greater than normal, DNA is not completely dissolved and thus magnetic beads may be adhered by viscous liquid. You need to add an appropriate volume of Buffer EB to completely dissolve it. If beads are still adhered, centrifuge the tube in a centrifuge at 8000 rpm for 1 minute.

4.3.2 Extracting the nucleic acids automatically on MGISP-NE384RS

4.3.2.1 Preparing consumables

According to the following table, prepare consumables for a workflow of automated extraction on MGISP-NE384RS and place them at room temperature until use:

Name	Brand	Cat. No.	Number
2.2 mL V-bottom 96-well deep-well plate	MGI	1000008088	20
96-well tips comb	MGI	1000025661	4

4.3.2.2 Preparing samples

You can extract 1 to 384 samples on MGISP-NE384RS.

Perform the following steps:

1. Take out a deep-well plate and mark it as 'sample'.
2. Perform different steps according to the sample type.
 - For samples of blood, amniotic fluid, cell and animal tissue, ensure that the sample is pretreated according to *Pretreating samples on Page 6* and add the sample into the plate for sample.
 - For other samples, add the sample and other reagents directly into the plate according to the following table.

Reagent name	Adding volume for each well (μL)		
	Whole blood/ fresh saliva	Blood of poultry, birds, or amphibians	Saliva (with saliva preservative)
Sample	200	V (5 to 10)	500
Buffer LS	/	200-V	/
Proteinase K	20	20	20
Buffer LB	300	300	/

3. Place the plate on ice until use.

4.3.2.3 Preparing reagents

Perform the following steps:


1. Add absolute ethanol into Buffer W1 according to the label.

2. Add absolute ethanol into Buffer W2 according to the label.
3. Take out 4 96-well deep-well plates. Add reagents according to the following table and mark the plates for each type of reagent.

Reagent name	Adding volume
Buffer W1	500 μ L/well
Buffer W2	600 μ L/well
Buffer W2	600 μ L/well
Buffer EB	60 μ L/well to 150 μ L/well

4.3.2.4 Starting extraction


Perform the following steps:

1. Switch to the  position to power on the device.
2. Turn on the computer and the desktop appears. Double-click the icon of MGISP-NE384RS to run the software.
3. Select **User** and **Real**, and enter the password. Click **Login** to enter the main interface.
4. Click **Initialize** on the top of the interface to start initializing.
You will be prompted after a successful initialization.
5. Empty the operation deck and close the door.
6. Select **Clean** in the main interface.
7. Click **Start**. The default duration is 20 minutes and you can also set the time as required.

UV lamps are turned on and air filter starts working.



CAUTION The ultraviolet radiation is harmful to the human body, so do not open the door after the cleaning starts.

8. Click **Process manage** >  to import the script.
9. Click  > **Workflow**. Click the drop-down list of **Script** and select **JB-A27-016 Genomic DNA Extraction Kit_RV1.0_SV1.0_EN**. Place samples, reagents and consumables according to the following table:

Reagent name	Position
Sample	Pos1
Buffer W1	Pos2
Buffer W2	Pos3

Reagent name	Position
Buffer W2	Pos4
Buffer EB	Pos6

10. Place 96-well tips comb according to the sample number.

11. Click **Run**. Select the required lanes and tips comb in the pop-up window. Click **OK**. The device starts extraction according to the following table. The whole workflow takes about 40 minutes.

	Step 1	Step 2	Step 3	Step 4	Step 5	Step 6	Step 7	Step 8
Step name	Lysis	Lysis	Bind	Wash	Wash	Wash	Elution	Release
Position	5	1	1	2	3	4	6	2
Volume (μL)	520	520	870	500	600	600	100	500
Delay time (s)	0	0	0	0	0	0	120	0
Mix	False	True	True	True	True	True	True	True
Mix type	/	Magnetic	Normal	Normal	Normal	Normal	Normal	Normal
Mix rate	Middle	Middle	Middle	High	High	High	High	High
Mix time (s)	1	900	180	180	120	120	300	5
Collect	True	False	True	True	True	True	True	False
Collect mode	Normal	Normal	Cycle	Cycle	Cycle	Cycle	Cycle	Normal
Collect cycle (time)	1	1	2	2	2	2	10	1
Collect time (s)	1	1	1	1	1	1	1	1
Dialog	False	True	False	False	False	False	False	False
Dialog content	/	Add 350 μL of isopropanol and 20 μL of Magnetic Beads H to each sample well of Pos1 plate.	/	/	/	/	/	/

Before step 3, you will be prompted to confirm that you have added 350 μ L of isopropanol and 20 μ L of Magetic Beads H to each sample well of Pos1 plate. Click **OK** and step 3 starts.

During the workflow, click **Pause** to pause and click **Resume** to resume the workflow if required.

The temperature control settings are as follows:

Position	Pos1	Pos6
Temperature	75 °C	56 °C
Open step	Step1	Step7
Close step	Step2	Step7
Action	Mix	Mix
Order	After	After

12. After the program ends, transfer the 96-well tips comb to the medical waste bag.
13. Immediately remove the 96-well plate from Pos6, seal the plate and store it in a freezer at -20 °C .

You can also transfer the DNA product in the 96-well plate from Pos6 to a new plate, seal and store it in a freezer at -20 °C .

Chapter 5 Warnings and precautions

- This product is for research use only. Please read the instructions for use carefully before use.
- Before experiment, be sure to be familiar with and master the operation methods and precautions of various devices to be used.
- You should prepare the isopropanol and RNase A (20 mg/mL) before the experiment.
- Please use recommended consumables for experiment.
- After the experiment, ensure that bottle caps, particularly caps of Buffer W1 and Buffer W2 containing the absolute ethanol, are securely tightened.
- Buffer EB consists of 10 mM Tris-HCl (pH 8.0) and 0.5 mM EDTA (pH 8.0). Please prepare the elution buffer according to your specific needs.
- Direct contact with skin and eyes should be avoided for all samples and reagents. Do not swallow. If accidental ingestion occurs, please get medical attention immediately. If skin exposure occurs, rinse with large amounts of water and get medical attention if irritation persists.

- All samples and wastes should be disposed of in accordance with relevant regulations.
- Do not use expired products.

Appendix 1 Manufacturer information

Manufacturer	Wuhan MGI Tech Co., Ltd.
Address	Building 24, Stage 3.1, BioLake Accelerator, No.388, 2nd Gaoxin Road, East Lake High-Tech Development Zone, 430075, Wuhan, P.R. China
	Building B13, No.818, Gaoxin Avenue, East Lake High-Tech Development Zone, 430075, Wuhan, P.R.China
Technical support	Wuhan MGI Tech Co., Ltd.
E-mail	MGI-service@mgi-tech.com
Website	en.mgi-tech.com